

TENDER NOTICE

Sealed Tender offers are invited from company/dealar only for supply of Eko milk ultra single sensor milk analyzers with Ultrasonic stirrer to our dairy. Approximate requirement of milk analyzers would be 25 nos. Terms & conditions & specifications of this analyzer is available at our office & on our website www.gokulmilk.coop Sealed Tender offers alongwith EMD of Rs. 10,000/- by D/D/RTGS issued in the name of Kolhapur Zilla Sahakari Dudh Utpadak Sangh Ltd., Kolhapur . Tender offer in a sealed cover duly marked as 'TENDER FOR EKO MILK ULTRA MILK ANALYZERS' should reach our office on or before 12.11.2025. Right to accept or reject any Tender offer without assigning reason is reserved.

Managing Director	Chairman
Kolhapur Zilla Sahakari Dudh Utpadak Sangh Ltd.	Kolhapur
B-1, M.I.D.C., Gokul Shirgaon, Kolhapur.	

General Terms & Conditions

- 1) Technical details of required Eko milk ultra single sensor milk analyzers with Ultrasonic stirrer is given in the attached annexure.
- 2) Tender offers is to be given on the Letter heads of the Tenderers in the format given by us.
- 3) Earnest Money Deposit of Rs. 10,000/- (Rs. Ten thousand only) by D/D/RTGS issued in favour of Kolhapur Zilla Sahakari Dudh Utpadak Sangh Ltd., Kolhapur must be given alongwith the Tender. Tender without EMD will not be considered.
- 4) Tender for Eko milk ultra single sensor milk analyzers with Ultrasonic stirrer submitted by main company or their authorised dealer only. Dealer must submitted authorisation certificate along with tender
- 5) Successful tenderer should give training regarding operation & repair of analyzer circuit diagram to our technician at free of cost.
- 6) Warranty – Company should give three years warranty of the analyzer machines installed by the company alongwith Off line application to change password of the machine. Comprehensive warranty including main PCB with sensor ,motor PCB ,motor & transformer from date of installation except consumable parts (Rubber, plastics etc.).
- 7) You have to submit price list of all required spares of milk analyzer.
- 8) Payment of the supplied analyzers will be made by Kolhapur Zilla Sahakari Dudh Utpadak Sangh Ltd. Kolhapur.
- 9) Sealed Tender offer duly marked as 'TENDER FOR EKO MILK ULTRA MILK ANALYZERS' is to be given at our Gokul Shirgaon, Kolhapur office on or before **-12.11.2025**
- 10) Company should provide analyzer machines and its parts of only original make of company. If it is found that you have provided duplicate machine or its parts then you will be penalized.
- 11) Within warranty period, if PCB is damaged by any reason within warranty period the company should replace new PCB free of cost.
- 12) Company should provide list of items which are not included in warranty period.
- 13) Company should attend and repair the machine within 24 hours after reporting the complaint. If the company is unable to repair the machine within 24 hours ,the company should provide standby machine till the machine is repaired.
- 14) While installation of new machine its calibration should be done in front of Sangh's technician at Sangh office situated at Tarabai Park or chilling centre or at BMC centre.
- 15) Within warranty period any damage is occurred, Company should rectify the milk analyzer at office situated at Tarabai Park, Chilling Centre or sub centre or BMC Center.
- 16) Company should maintain sufficient service staff to attend the complaints within our milk collection areas.
- 17) The Sangh will change passwords of the machines provided to the Milk Societies. Company should not tamper/ modify directly the passwords given by the sangh. If such incident is found, you will be penalised.
- 18) Right to accept or rejects any or all Tenders without assigning any reason is reserved.
- 19) Subject to Kolhapur Jurisdiction.

Annexure
Specifications of Eko milk ultra single sensor milk analyzers with Ultrasonic stirrer

Sr. No.	Measuring Parameters	
	Parameter	
1	Fat	Cow
		Buffalo
2	SNF (Solid non fat)	Cow
		Buffalo
3	Density (CLR)	
4	Added Water	
5	Protein	
6	Lactose	

Electrical Parameters		
1	AC Power with cable	230 v/ 50 Hz
2	DC power with battery cable	12 v to 14 v

Interface	
1	RS 232 Serial port & USB port with 2 meter cable

Ultrasonic Stirrer		
1	AC Power with cable	230 v/ 50 Hz
2	DC power with battery cable	12 v to 14 v

Format

TENDER OFFER

To,

The Managing Director
Kolhapur Zilla Sahakari Dudh
Utpadak Sangh Ltd.,
B-1, M.I.D.C., Kolhapur.

Sub - Tender offer for Eko milk ultra single sensor milk analyzers
with Ultrasonic stirrer

Dear Sir,

In continuation to your Tender Notice published in we are submitting our Tender offer
for the supply of Eko milk ultra single sensor milk analyzers with Ultrasonic stirrer our main
dairy.

We have studied Terms & conditions of Tender and we agree for these terms - conditions .

Yours faithfully

Seal & Signature of Tenderer

Encl. -

- 1) Price details
- 2) DD/RTGS towards EMD

PRICE DETAILS

Sr. No.	Particulars of Analyzer	Rate per machine
1)	Unit Price of Eko milk ultra single sensor milk analyzers with Ultrasonic stirrer (Including Transport & Installation charges)	
2)	Applicable GST	
	Total Unit Price	

Other information if any -

**Seal & Signature of
Tenderer**